




America

CERTIFICATE

No. U8 094770 0002 Rev. 01

Holder of Certificate: **Tornado Spectral Systems Inc.**
5155 Spectrum Way, Unit 6
Mississauga ON L4W 5A1
CANADA

Certification Mark:

Product: **Laboratory equipment**
Model(s): **HyperFlux PRO PLUS 785**
Parameters: Rated Input Voltage: 120/240 V~
Rated Frequency: 50/60 Hz
Rated Input Current: 1.6 A
Protection Class: I

Tested according to: CAN/CSA C22.2 No. 61010-1:2012
UL 61010-1:2012

This product was voluntarily tested to the relevant safety requirements referenced on this certificate. It can be marked with the certification mark above. The mark must not be altered in any way. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC 17067. Certification is based on the TÜV SÜD "Testing and Certification Regulations". TÜV SÜD America Inc. is an OSHA recognized NRTL and a Standards Council of Canada accredited Certification body.

Test report no.: 7169000118-001
Valid until: 2026-05-10
Date, 2021-06-03

(Vivek Chandna)