




America

CERTIFICATE

No. U8 094770 0003 Rev. 01

Holder of Certificate: **Tornado Spectral Systems Inc.**
 5155 Spectrum Way, Unit 6
 Mississauga ON L4W 5A1
 CANADA

Certification Mark:



Product: **Laboratory equipment**
Model(s): **OPIS 35**

Parameters:

Rated Input Voltage:	120/240V~
Rated Frequency:	50/60Hz
Rated Input Current:	1.6A
Protection Class:	I
Laser Class:	Class 3B

Tested according to: CAN/CSA-C22.2 No. 61010-1/2012
 UL 61010-1:2012

This product was voluntarily tested to the relevant safety requirements referenced on this certificate. It can be marked with the certification mark above. The mark must not be altered in any way. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC 17067. Certification is based on the TÜV SÜD "Testing and Certification Regulations". TÜV SÜD America Inc. is an OSHA recognized NRTL and a Standards Council of Canada accredited Certification body.

Test report no.: 7169005599-001

Date, 2021-06-03

(Vivek Chandna)