




America

CERTIFICATE

No. U8 094770 0004 Rev. 00

Holder of Certificate: **Tornado Spectral Systems Inc.**
 5155 Spectrum Way, Unit 6
 Mississauga ON L4W 5A1
 CANADA

Certification Mark:



Product: **Laboratory equipment**
Process Guardian Raman

Model(s): **PGRxxxx**
 Where 'xxxx' is any alphanumeric characters not affecting safety.

Parameters: Rated Input Voltage: 120/240 V~
 Rated Frequency: 50/60 Hz
 Rated Input Current: 1.1 A
 Protection Class: I
 Laser Class: 3B

Tested according to: CAN/CSA-C22.2 No. 61010-1:2012/R:2019-07
 UL 61010-1:2012/R:2019-07

This product was voluntarily tested to the relevant safety requirements referenced on this certificate. It can be marked with the certification mark above. The mark must not be altered in any way. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC 17067. Certification is based on the TÜV SÜD "Testing and Certification Regulations". TÜV SÜD America Inc. is an OSHA recognized NRTL and a Standards Council of Canada accredited Certification body.

Test report no.: 7169010267-000

Date, 2021-12-17

(Peter Keith)