

DOC: QC-132

REV: 0

**REVIEWED BY:** 

Riyan Mohiyuddin

APPROVED BY:

Scott Baker

Apr 28, 2023

## **UK CA DECLARATION OF CONFORMITY**

Manufacturer: Tornado Spectral Systems Inc.

5155 Spectrum Way, Unit 6 Mississauga, ON, L4W 5A1

Tel: 416-361-3444 Fax: 416-361-6374 operations@tornado-spectral.com

UK Authorized **Obelis UK Ltd.** 

Representative: Sandford Gate, East Point Business Park

Oxford, OX4 6LB, United Kingdom

Tel: +44.1491.378012 Email: info@obelis.co.uk

Manufacturer declares under its sole responsibility that the product (Raman Spectrometer Instrument)

<u>HyperFlux PRO Plus 785 Product</u> Serial Numbers: <u>HFPP 0050 to HFPP 9999</u>

The object of declaration is in compliance with the relevant statutory requirements:

The Electromagnetic Compatibility Regulations 2016 (as amended)

Electrical Equipment (Safety) Regulations 2016 (as amended)

The Restriction of the Use of Certain Hazardous Substances in Electrical and Electronic Equipment

Regulations 2012 (as amended)

The designated standards or other technical specifications to which conformity is declared:

LVD: EN 61010-1:2010-10 /A1:2019

EMC: EN 61326-1:2013 EN 61000-6-2:2005

EN 61000-6-4:2007 + A1:2011

This statement is based on a single evaluation of one sample of the above-mentioned product. It does not imply an assessment of all serial numbers within the product line. The manufacturer shall ensure that all products in this product line are in conformity with the samples detailed in this report.

The last two digits of the year in which the UK CA marking was affixed: 23.

Signed for and behalf of Tornado Spectral Systems Inc.

Place	Signature	Date
Mississauga, Ontario	Riyan Mohiyuddu	28-Apr-2023

Name and Position: Riyan Mohiyuddin, Quality Manager

5155 Spectrum Way, Unit 6

Mississauga, ON, L4W 5A1, Canada.